

OLYMPUS[®]

Your Vision, Our Future

EVIS EXERA II VIDEO SYSTEM CENTER

OLYMPUS CV-180

EVIS
EXERA II

*Experience the new standard in
endoscopic imaging with this advanced
HDTV-compatible video processor*



Main Features

- Equipped with high-resolution HDTV and Narrow Band Imaging capabilities to provide the best possible image quality for endoscopes and laparoscopes, enhancing observation of capillaries and mucosal tissues.
- Compatible with EVIS 100/130/140 Series and EVIS EXERA 160 Series endoscopes in addition to EVIS EXERA II 180 Series models as well as bronchoscopes and surgical endoscopes (EndoEYE videoscopes, flexible VISERA videoscopes, and VISERA 1CCD/OTV-SP1 3CCD camera heads).
- NBI (Narrow Band Imaging) to enhance the visibility of capillaries and other structures on the mucosal surface (only for GI).
- Two types of structure enhancement are available — the original Type A for observation of larger mucosal structures with high contrast and the new Type B for observation of smaller structures, such as capillaries.
- Electronic magnification enlarges moving images at the touch of a button on the scope or on the keyboard by 1.2X or 1.5X.
- HD/SD SDI output for high-quality video image transfer.
- Convenient digital-to-digital recording of both the still and moving images. Still images are stored on xD cards via the PC card adapter, and moving images are stored on digital video recorders via IEEE 1394 (DV interface) using remote control switches.
- Automatic Iris eliminates the need for switching between Peak and Average required in conventional manual adjustment.
- Picture-in-picture (PiP) display for any combination of endoscopic images, fluoroscopic images, ultrasound images, laparoscopic images, and images from the endoscope position-detecting unit.
- Convenient index display for documentation.
- Scope ID function for easier endoscopy suite management and for next-generation system expansion.



Specifications

Observations	HDTV signal output	Either RGB or YPbPr output can be selected.			
	SDTV signal output	VBS composite (NTSC), Y/C and RGB; simultaneous output possible.			
	White balance adjustment	White balance adjustment is possible using the white balance button on the front panel.			
	Standard color chart output	A color bar chart can be displayed by pressing the "SHIFT" + "F7" keys on the keyboard.			
	Color tone adjustment	The following color tone adjustments are possible using the "COLOR" key and "arrow" keys on the keyboard: • Red adjustment: +/-8 steps • Blue adjustment: +/-8 steps • Chroma adjustment: +/-8 steps			
	Automatic gain control (AGC)	The image can be electrically amplified when the light is inadequate due to the distal end of the endoscope being too far from the object.			
	Contrast	The image contrast can be set to one of the following three modes (N, H, L) using the "SHIFT" + "F6" key on the keyboard. • N (Normal): Normal image. • H (High): The dark areas are darker and the bright areas are brighter than in the normal image. • L (Low): The dark areas are brighter and bright areas are darker than in the normal image.			
	Iris	The auto iris modes can be selected using the "iris mode" button on the front panel. • Peak: For use when observing by focusing on a small bright area. • Auto: For use when observing by focusing on the image center.			
	Image enhancement setting	Fine patterns or edges in the endoscopic images can be enhanced electrically to increase the image sharpness. Either the structural enhancement or edge enhancement can be selected according to the user setup.			
	Switching the enhancement modes	The enhancement level can be selected from four levels (OFF, 1, 2 and 3) using the image enhancement switch on the front panel.			
	Image size selection	The size of the endoscopic image can be changed using the "SHIFT" + "F8" key on the keyboard.			
	Reset to defaults	The following settings can be reset to their defaults using the reset button on the front panel: • User preset • Image source • Color tone • Freeze • Release index • Zoom • Special light observation • Arrow pointer • Stopwatch • Characters on screen • Exposure • PiP			
	Freeze	An endoscopic image is frozen using an endoscope or "FREEZE" key on the keyboard.			
	Documentation	Remote control	The following ancillary equipment can be controlled from the front panel, keyboard or endoscope's remote switches: (Specified models only) • VCR • Video printer • Image filing system • Endoscopic ultrasound center, etc.		
		Patient data	The following data can be displayed on the monitor by keyboard: • Patient ID No. • Patient name • Sex & age • Date of birth • Date of recording (time, stopwatch) • Image frame No. • Videotape recorder mode • Display image setting • Physician name • Comments		
Advance registration of patient data		The following data of up to 40 patients can be entered prior to surgery using the keyboard: • Patient ID No. • Patient name • Sex & age • Date of birth • Physician			
PC card	Media	xD-Picture Card (1G/512/256/128/64/32/16 MB), specified by Olympus. MAPC-10 can be used as PC card adapter.			
	Recording format	TIFF: no compression, SHQ: approx. 1/5, HQ: approx. 1/7, SQ: approx. 1/10.			
	Number of recording images	In 16 MB, SDTV/HDTV, TIFF: approx. 30/6 images, SHQ: approx. 310/110 images, HQ: approx. 2000/760 images, SQ: approx. 2570/430 images.			
Image storage and retrieval	Monitor output	Using the monitor out switches on the front panel, it is possible to select an image from the endoscope or ancillary equipment for display on the monitor.			
Memory backup	Memorization of selected setting	The following settings are held in memory even after the video system center is turned OFF: • White balance • Iris mode • Enhancement • Image size • Color tone			
Classification as medical electrical equipment	Type of protection against electric shock	Class I			
	Degree of protection against electric shock of applied part	Depends on applied part. See also applied part for camera head or videoscope.			
	Degree of protection against explosion	The video system center should be kept away from flammable gases.			
Power supply	Voltage (Voltage fluctuation)	100-240 VAC (Within ±10%)			
	Frequency (Frequency fluctuation)	50/60 Hz (Within ±1%)	Consumption electric power	150 VA	
Size	Fuse rating	5 A, 250 V	Fuse size	5 × 20 mm	
	Dimensions	382 (W) × 91 (H) × 490 (D) mm (Maximum)	Weight	10.5 kg	

Specifications, design and accessories are subject to change without any notice or obligation on the part of the manufacturer.



OLYMPUS MEDICAL SYSTEMS CORP.
Shinjuku Mondolith, 3-1 Nishi-Shinjuku 2-chome, Shinjuku-ku, Tokyo 163-0914, Japan
OLYMPUS WINTER & IBE GMBH
Kuehnstrasse 61, 22045 Hamburg, Germany
OLYMPUS MEDICAL SYSTEMS EUROPA GMBH
Postfach 10 49 005, 20034 Hamburg / Wendenstrasse 14-18, 20097 Hamburg, Germany
OLYMPUS SURGICAL & INDUSTRIAL AMERICA INC.
One Corporate Drive, Orangeburg, New York 10962, U.S.A.
OLYMPUS AMERICA INC.
3500 Corporate Parkway, P.O. Box 610, Center Valley, PA 18034-0610, U.S.A.
OLYMPUS LATIN AMERICA, INC.
5301 Blue Lagoon Drive, Suite 290 Miami, Florida 33126-2097, U.S.A.
KEYMED LTD.
Keymed House, Stock Road, Southend-on-Sea, Essex SS2 5QH, England

OLYMPUS SINGAPORE PTE LTD.
491B River Valley Road #12-01/04, Valley Point Office Tower, Singapore 248373
OLYMPUS HONG KONG AND CHINA LIMITED
Room 1520-1527, Ocean Centre, 5 Canton Road, Tsimshatsui, Kowloon, Hong Kong
OLYMPUS (BEIJING) SALES & SERVICE CO., LTD.
Room 1202, NO.1 Tower, A12 Jianguomenwai Avenue Chaoyang District Beijing, 100022, China
OLYMPUS MOSCOW LIMITED LIABILITY COMPANY
117071, Moscow, Malaya Kaluzhskaya 19, bld. 1, 1L2, Russia
OLYMPUS AUSTRALIA PTY LTD
31 Gilby Road, Mount Waverley, VIC., 3149, Australia